

使用測定器: CK 1620(NF製)

\* 2/3,3/3のTTLは入力直近でターミネートをして測定

		JOB.10609(1/3) 最大周波数 (MHz)	JOB.10609(2/3) 最大周波数 (MHz)	JOB.10609(3/3) 最大周波数 (MHz)
TTL入力 (0~3.3V)	CH0	190	240	240
	CH1	190	210	200
	CH2	190	200	200
	CH3	190	200	200
	CH4	190	200	200
	CH5	190	200	200
	CH6	190	240	240
	CH7	180	240	240
NIM入力 (0~-1V)	CH0	340	340	330
	CH1	380	390	370
	CH2	440	460	430
	CH3	330	340	330
	CH4	330	330	350
	CH5	400	390	400
	CH6	460	440	480
	CH7	460	410	450

